



Medicines Authority
Inspectorate and Enforcement Division
203, Level 3, Rue d'Argens,
Gzira
Malta

Telephone: +356 2343 9142

Facsimile: +356 2343 9162

Drug Alert

CLASS 3 MEDICINES RECALL

Action Within 5 Days

Date: 13th October 2010

Our Ref: MDR 3-10/10

Dear Healthcare Professional,

Poligas Ltd

**Worthington Medical Oxygen cylinders 35L, 40L, 50L
(liquid volume)**

N/A

Batch Number	Expiry Date	Pack Size	First Distributed
All Worthington Medical Oxygen cylinders 35L, 40L & 50L (see corresponding list for cylinder numbers)	N/A	N/A	N/A

Following the receipt of two complaints regarding various Medical Oxygen cylinders the Inspectorate and Enforcement Directorate within the Medicines Authority requested a precautionary recall of the affected cylinders.

The recall has been classified as a Class III recall and is being carried out by Poligas Ltd. (Marketing Authorisation Holder and Manufacturer) under the supervision of the Medicines Authority.

The reason for the precautionary recall of all cylinders is the presence of an odour in a number of Medical Oxygen gas cylinders

Thanks and regards

Karl De Marco
Medicines Inspector