

To the attention of Medical Device Vigilance  
responsible / Central Pharmacy

Saint Priest, May 24<sup>th</sup>, 2024

---

**Subject: URGENT - FIELD SAFETY NOTICE – INTEGRA – CUSA® Excel 23KHz Straight Handpiece – Reference: C2600 – FIELD SAFETY ACTION**

**Legal manufacturer:** Integra Lifesciences, IDA Business and Technology Park, Sragh Ave, Tullamore, Co. Offaly – IRELAND. IE-MF-000003849

**Medical device:**

The CUSA® Handpiece is intended for use with the CUSA® Excel/CUSA® Excel+ Ultrasonic Surgical Aspirator System and accessories. The CUSA® Excel/ CUSA® Excel+ Ultrasonic Surgical Aspirator System is an ultrasonic tissue ablation system that allows a surgeon to remove tissue selectively. It performs three functions:

1. Fragmentation: when the vibrating tip of the handpiece comes into contact with tissue, the cells of the tissue break apart or “fragment”.
2. Irrigation: irrigation fluid from a user-supplied saline bag or Lactated Ringer’s solution is transferred to the distal tip of the handpiece.
3. Aspiration (Suction): draws or “aspirates” irrigation fluid, fragmented tissue and other material through the distal end of the surgical tip to the user-supplied canister.

**Primary clinical purpose of device:**

The CUSA® Excel/CUSA® Excel+ Ultrasonic Surgical Aspirator System is indicated for use in these surgical procedures where fragmentation, emulsification and aspiration of soft and hard tissue is desirable:

- Neurosurgery
- Gastrointestinal and affiliated organ surgery
- Urological surgery
- General surgery
- Orthopedic surgery
- Gynecological surgery
- Laparoscopic surgery

**Concerned reference:**

C2600

**Serial Numbers:**

All serial numbers distributed from 1<sup>st</sup> of January 2017 to 14<sup>th</sup> of December 2023 (*refer to paragraph B. How to confirm the serial number*)

---



Dear Valued Integra Customer,

The purpose of this letter is to notify you that Integra LifeSciences is issuing a voluntary medical device correction for any serial numbers of the CUSA Excel Handpiece manufactured prior to November 2023.

The decision to conduct a voluntary correction of the product was based on the following: through an internal investigation of customer complaints, it was identified that there is a potential for the CUSA Excel 23KHz Handpiece housing to crack. In order to mitigate this issue, the devices will be reworked through our service and repair process for the units manufactured prior to November 2023.

**Note: A replacement housing was introduced in November 2023 to fix the crack housing, as such any new parts manufactured after November 2023 do not have the potential of cracked housing issue. These devices started getting distributed on December 12<sup>th</sup>, 2023.**

**Table 1: Impacted Product Information**

Manufacturer's Product Number (Catalog Number)	Product Name (Description)	UDI Number	Serial Number	Expiration Date	Distribution Dates (DD/M/YYYY)
C2600*	CUSA® Excel 23KHz Straight Handpiece	10381780039419	All Serial Numbers Distributed before 14/12/2023	N/A	Before 14/12/2023

\*Note: you may see C2600P or C2600SRL on labeling. This notice is applicable for those SKUs as well.

This voluntary correction is limited to the reference and specific serial numbers outlined in Table 1.

**Risk to Health**

Based on the health hazard evaluation conducted for this issue, the potential harms are listed in the following table:

Hazardous Situation	Harm
System does not function during use	Inconvenience to User (no or minor delay in patient treatment)
System does not function prior to use	Inconvenience to User
System does not function as intended during use	Delay in patient treatment within the standard of care (e.g. leading to increased time under anesthesia, less than 30 minutes) may result in transient post-operative confusion)
	Clinically relevant delay in treatment beyond the standard of care (e.g. greater than 30 minutes); older individuals may be at increased risk of postoperative fever

Note: Prolonged anesthesia time risks may vary among different patient populations i.e., age, underlying health conditions, type of surgery, and individual responses to anesthesia. However, elderly are more susceptible to risks of postoperative confusion and fever.

There have been a total of 238 complaints reported worldwide as of April 15, 2024, due to this issue. Out of the 238 worldwide complaints, 6 serious incidents with clinically relevant delay in treatment beyond the standard of care (e.g. greater than 30 minutes); have been reported for Europe.

**Actions to be Taken by Customers:**

- A. Please **review and understand** the information provided in this letter.
- B. Confirm the serial number of the CUSA Excel 23KHz straight handpiece as follows:

All serial numbers for the CUSA Excel 23KHz straight handpiece start with:

- i. HA – indicating the product family followed by
- ii. Letter – indicating the month of manufacture where A = January; B = February; C = March etc. Noting that the letter I is not used.
- iii. Two digits e.g 24 indicating the year of manufacture with 24 representing 2024.
- iv. The next set of 3 digits are random serial numbers unique to each hand piece.
- v. The last 2 digits represent the model number.
- vi. IE indicates the location of manufacture.

Example: serial number HAL2300103IE represents a product that is manufactured in November 2023 and is acceptable for continued use.

- C. If your handpiece was manufactured prior to HAL23XXXXXXIE you should take the following actions:
  - 1. Inspect the housing for the presence of a crack – see image below.

Figure 1: Example of Cracked Housing

Good Part	Bad Part
Reported failure not applicable C2600 housing	Reported failures: housing crack/ cracked, damaged, C2600 housing
	

- 2. The Handpiece cracked housing may be visually detected by clinical staff during inspections prior to or post-surgery. If a crack is noted, please remove the handpiece and return to Integra Service. TCS Benelux TCS.Benelux@integralife.com; TCS DACH TCS.DACH@integralife.com; TCS France TCS.France@integralife.com; TCS Italy TCS.Italy@integralife.com; (*will be added to the corresponding FSN depending on the country*). Check the box "I do have affected units" in the enclosed the reply form.
- 3. If a crack is not noted, Regardless of the service contract status, the housing will be replaced free of charge during your next service.

**\*\*Please note that as per our standard IFU the recommended service frequency, should be performed every 50 hours of use or 100 procedures whichever comes first\*\*.**

- D. Please return the completed reply form by email to [emea-fsca-neuro@integralife.com](mailto:emea-fsca-neuro@integralife.com), or Fax to +33 (0)4.37.47. 59.30.



By filling in this form, you confirm that you have received this Safety Notice and you intend to fully comply with this notification. **We expect a response within 3 weeks.** You also confirm that this notification has been forwarded to every concerned person in your organization.

We recommend that you retain a copy of the form for your records.

The receipt of this form ensures that Integra has achieved a level of effectiveness in communicating this information.

National Competent Authorities may perform audits of field actions of this nature to verify that our customers have been notified and understand the nature of the field action being taken.

The National Competent Authority of your country has been alerted of this Field Safety Corrective Action.

Thank you for your cooperation with this Field Safety Corrective Action and for returning the attached Reply Form.

Please feel free to contact our Post Market Surveillance Department at [emea-fsca-neuro@integralife.com](mailto:emea-fsca-neuro@integralife.com) for any additional questions. Your cooperation is appreciated, and we thank you for your continued collaboration.

Yours Sincerely,

Post Marketing Surveillance Department

**Appendix:** Field Safety Notice Distributor Reply Form (2 pages)

# CUSTOMER REPLY FORM

1. Field Safety Notice (FSN) information	
FSN Reference number	FSN-2024-HHE-006
FSN Date	24/05/2024
Device name	CUSA® Excel 23KHz Straight Handpiece
Product Code	C2600
Serial Numbers	All serial numbers <u>prior</u> to HAL23XXXXXXIE (distributed before 14/12/2023)

2. Customer Details	
Account Number	
Healthcare Organisation Name*	
Organisation Address*	
Department/Unit	
Shipping address if different to above	
Contact Name*	
Title or Function	
Telephone number*	
Email*	

3. Customer action undertaken on behalf of Healthcare Organisation		
<input type="checkbox"/>	I confirm receipt of the Field Safety Notice and that I read and understood its content.*	
<input type="checkbox"/>	I performed all actions requested by the FSN *	
<input type="checkbox"/>	The information and required actions have been brought to the attention of all relevant users and executed. *	
<input type="checkbox"/>	I have checked my inventory*	
<input type="checkbox"/>	I <u>do have</u> affected units	<b>Quantity:</b> <b>Serial numbers:</b>
<input type="checkbox"/>	I <u>do not</u> have any affected units	
<input type="checkbox"/>	I have a query please contact me	<i>Customer to enter contact details if different from above and brief description of query</i>
	Print Name*	<i>Customer print name here</i>
	Signature*	<i>Customer sign here</i>
	Date*	

<b>4. Return acknowledgement to Sender</b>	
Email	<a href="mailto:emea-fsca-neuro@integralife.com">emea-fsca-neuro@integralife.com</a>
Distributor Helpline	+33 (0) 6 30 20 69 66
Postal Address	Post Market Surveillance Department Integra Immeuble Séquoia 2, 97 allée Alexandre Borodine Parc technologique de la Porte des Alpes 69800 Saint Priest, France
Web Portal	<a href="https://integralife.eu/">https://integralife.eu/</a>
Fax	+33 (0)4 37 47 59 30
Deadline for returning the customer reply form*	14/06/2024

Mandatory fields are marked with \*

It is important that your organisation takes the actions detailed in the FSN and confirms that you have received the FSN.

Your organisation's reply is the evidence we need to monitor the progress of the corrective action.