

## Country-Specific Glucose Monitor List

**Country Name:** Malta

### **Important Information**

This is a non-comprehensive list, current as of **23 July 2024**. Absence of a specific glucose monitor or test strips from this list does NOT imply compatibility or incompatibility with EXTRANEAL (Icodextrin) PD solution. Similarly, other glucose-measuring technologies which are not listed below (such as continuous glucose monitoring systems) may or may not be compatible with EXTRANEAL. **Always contact the device manufacturer for current information.** If the manufacturer cannot provide information regarding compatibility of the device with Icodextrin and Maltose, Baxter does NOT recommend that EXTRANEAL patients use the product. Baxter reserves the right to change this list without notice and does not represent that it includes all potentially incompatible products.

The glucose monitors manufacturers listed have certified to Baxter that they have tested (as per ISO 15197) their monitors with Maltose and Icodextrin to Baxter's recommended limits: 278 mg/dL (Maltose) and 1094 mg/dL (Icodextrin). The manufacturers certified that their "green" monitors mentioned below showed no interference of blood glucose readings. Please note that the compatibility list below is only for brand-name monitors used with the corresponding brand-name test strip. If a brand-name monitor is used with another manufacture's test strip, **always contact the test strip manufacturer for current information.**

This list is compiled from a search via: Internet, Literature, Baxter internal studies, information from government agencies, test strip leaflets, safety alerts, and direct information from the product manufacturers. While efforts have been made to provide accurate and current information, Baxter does not manufacture these glucose monitors or test strips and thus does not guarantee the initial or continued accuracy of this information. Please contact the manufacturer(s) of the glucose monitor and test strip to obtain the latest compatibility information before using in conjunction with EXTRANEAL (Icodextrin) Peritoneal Dialysis (PD) solution.

1. EXTRANEAL (Icodextrin) PD solution contains Icodextrin. Maltose, a metabolite of Icodextrin, may interfere with certain glucose monitors or test strips. This interference will result in a falsely elevated glucose reading using these monitors or test strips.
2. This interference may mask true hypoglycemia or lead to the erroneous diagnosis of hyperglycemia. Thus, a blood glucose reading within or above the normal range in a patient on EXTRANEAL (Icodextrin) PD solution, using these monitors or test strips, may mask true low blood sugar. This may cause a patient or health care professional to not take the appropriate steps to bring the blood sugar into a normal range. Alternatively, a falsely elevated blood glucose reading could cause a patient to get more insulin than needed. Both of these situations can lead to life-threatening events, including loss of consciousness, coma, neurological damage or death.
3. The dialysis unit or patient should contact the manufacturer of the glucose monitor and/or test strips to determine if the monitor or test strips they are using are subject to interference by Icodextrin or Maltose. Also, consult the product information included with the glucose monitor and test strips.
4. The following list is for reference only. This list does not imply recommendation of these glucose monitors or test strips.
5. Identified compatibilities are shown in the table below. **ONLY** glucose-specific monitors and test strips should be used with patients on EXTRANEAL (Icodextrin) PD solution. Contact the manufacturer to verify that the test strips and monitor are glucose specific. The list provides the contact information of the more common, major brand manufacturers.

### **Reporting of suspected adverse reactions to the Malta Medicines Authority**

Reporting suspected adverse reactions after authorisation of the medicinal product is important. It allows continued monitoring of the benefit/risk balance of the medicinal product. Healthcare professionals are asked to report any suspected adverse reactions via ADR Reporting Website: [www.medicinesauthority.gov.mt/adrportal](http://www.medicinesauthority.gov.mt/adrportal)

Alternatively, any suspected adverse reactions can also be reported to Malta Medicines Authority Post-Licensing Directorate, Sir Temi Zammit Buildings, Malta Life Sciences Park, San Ġwann SGN 3000, Malta

**Reporting of suspected adverse reactions to the Marketing Authorisation Holder**

Suspected adverse reactions to Baxter products can also be reported direct to Baxter Pharmacovigilance on +44 (0)1635 206360, or by email to [vigilanceuk@baxter.com](mailto:vigilanceuk@baxter.com)

**COUNTRY-SPECIFIC GLUCOSE MONITORS**

Updated 23 July 2024

Glucose Monitor Brand	Compatible with Extraneal (Icodextrin) PD solution (Glucose-specific)	Test Type*	Manufacturer
FreeStyle Freedom	Yes	GDH-FAD	Abbott Diabetes Care <a href="http://www.abbottdiabetescare.com">www.abbottdiabetescare.com</a> Phone: +356 2145 4166
FreeStyle Freedom Lite	Yes	GDH-FAD	
FreeStyle Lite	Yes	GDH-FAD	
FreeStyle Libre <sup>1</sup>	Not Recommended	GO	
FreeStyle Libre Pro <sup>1</sup>	Not Recommended	GO	
FreeStyle Libre 2 <sup>1</sup>	Not Recommended	GO	
FreeStyle Libre 3 <sup>1</sup>	Not Recommended	GO	
FreeStyle Optium Neo	Yes	GDH-NAD	
FreeStyle Optium Neo H	Yes	GDH-NAD	
FreeStyle Papillon Vision	Yes	GDH-FAD	
FreeStyle Precision Neo	Yes	GDH-NAD	
FreeStyle Precision Pro	Yes	GDH-NAD	
Optium Xido Neo	Yes	GDH-NAD	
Precision Xceed Pro	Yes	GDH-NAD	
Assure Platinum	Yes	GO	Arkray, Inc. <a href="http://www.arkray.co.jp">www.arkray.co.jp</a> Phone: N/A
GLUCOCARD 01	Yes	GO	
GLUCOCARD 01-mini	Yes	GO	
GLUCOCARD 01-mini plus	Yes	GO	
GLUCOCARD MX	Yes	GDH-FAD	
GLUCOCARD S	Yes	GDH-FAD	
GLUCOCARD SM	Yes	GDH-FAD	
GLUCOCARD Vital	Yes	GO	
GLUCOCARD X-meter	Yes	GDH-FAD	
GLUCOCARD X-mini	Yes	GDH-FAD	
GLUCOCARD X-mini plus	Yes	GDH-FAD	
GLUCOCARD Σ	Yes	GO	
GLUCOCARD Σ-mini	Yes	GO	
ReliOn Confirm	Yes	GO	
ReliOn micro	Yes	GO	
ReliOn Prime	Yes	GO	
GLUCOCARD S onyx	Yes	GDH-FAD	
PocketChem BG	Yes	GDH-FAD	
GLUCOCARD W	Yes	GO	
GLUCOCARD PlusCare	Yes	GDH-FAD	

Glucose Monitor Brand	Compatible with Extraneal (Icodextrin) PD solution (Glucose-specific)	Test Type*	Manufacturer
<b>GLUCOCARD Prime</b> <b>GlucoCard G+</b> <b>GlucoCard G+ Link</b> <b>GlucoCard <math>\Sigma</math>+</b> <b>GlucoCard <math>\Sigma</math>+ Link</b>	<b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b>	<b>GDH-FAD</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b>	
<b>Breeze 2</b> <b>Contour</b> <b>Contour Link</b> <b>Contour Next</b> <b>Contour Next EZ</b> <b>Contour Next Link</b> <b>Contour Next Link 2.4</b> <b>Contour Next USB</b> <b>Contour Next One</b> <b>Contour Plus</b> <b>Contour Plus Link 2.4</b> <b>Contour Plus One</b> <b>Contour TS</b> <b>Contour USB</b> <b>Contour XT</b> <b>Contour Next (Connected)</b> <b>Contour Care</b> <b>Contour Plus Blue</b> <b>Contour Plus Elite</b> <b>Contour Next GEN</b>	<b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b>	<b>GO</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b>	<b>Ascensia Diabetes Care</b> (formerly Bayer Healthcare) <a href="http://www.ascensia.com">www.ascensia.com</a> <b>Phone: N/A</b>
<b>OneTouch SelectPlus</b> <b>OneTouch SelectPlus Flex</b> <b>OneTouch SelectPlus Simple</b> <b>OneTouch Select Simple</b> <b>OneTouch Ultra 2</b> <b>OneTouch UltraEasy</b> <b>OneTouch UltraMini</b> <b>OneTouch UltraPlus Flex</b> <b>OneTouch UltraPlus Reflect</b> <b>OneTouch UltraVue</b> <b>OneTouch Verio</b> <b>OneTouch VerioFlex</b> <b>OneTouch VerioIQ</b> <b>OneTouch VerioPro+</b> <b>OneTouch VerioVue</b> <b>OneTouch Verio Reflect</b>	<b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b>	<b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GO</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b>	<b>Lifescan, Inc.</b> <a href="http://www.lifescan.com">www.lifescan.com</a> <b>Phone: +356 2124 201</b>

Glucose Monitor Brand	Compatible with Extraneal (Icodextrin) PD solution (Glucose-specific)	Test Type*	Manufacturer
<b>Nova Max Plus</b> <b>Nova Max Link</b> <b>StatStrip Hospital Glucose Meter</b> <b>StatStrip Hospital Glucose and Ketone Monitoring System</b> <b>StatStrip Xpress Glucose Meter</b> <b>StatStrip Xpress Glucose and Ketone Monitoring System</b> <b>StatStrip Xpress2 Glucose and Ketone Monitoring System</b> <b>Nova Pro Glucose/Ketone Meter</b> <b>StatStrip Glucose Hospital Meter System V2.0</b> <b>StatStrip Glucose/Ketone Hospital Meter System V2.0</b>	<b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b>  <b>Yes</b> <b>Yes</b>  <b>Yes</b> <b>Yes</b>  <b>Yes</b>	<b>GO</b> <b>GO</b> <b>GO</b> <b>GO</b>  <b>GO</b> <b>GO</b>     <b>GDH-FAD</b> <b>GO</b>  <b>GO</b>	<b>Nova Biomedical</b> <a href="http://www.novabiomedical.com">www.novabiomedical.com</a> <b>Phone: N/A</b>
<b>Accu-Chek Active</b> <b>Accu-Chek Aviva</b> <b>Accu-Chek Aviva Combo</b> <b>Accu-Chek Aviva Connect</b> <b>Accu-Chek Aviva Expert</b> <b>Accu-Chek Aviva Insight</b> <b>Accu-Chek Aviva Plus<sup>2</sup></b> <b>Accu-Chek Aviva Solo</b> <b>Accu-Chek Guide</b> <b>Accu-Chek Guide Me</b> <b>Accu-Chek Inform II</b> <b>Accu-Chek Instant</b> <b>Accu-Chek Instant S</b> <b>Accu-Chek Nano<sup>2</sup></b> <b>Accu-Chek Nano SmartView<sup>2</sup></b> <b>Accu-Chek Mobile</b> <b>Accu-Chek Performa</b> <b>Accu-Chek Performa Combo</b> <b>Accu-Chek Performa Connect</b> <b>Accu-Chek Performa Insight</b> <b>Accu-Chek Performa Solo</b> <b>Accu-Chek Guide Solo</b>	<b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b> <b>Yes</b>	<b>Mut Q-GDH</b> <b>Mut Q-GDH</b> <b>Mut Q-GDH</b> <b>Mut Q-GDH</b> <b>Mut Q-GDH</b> <b>Mut Q-GDH</b> <b>Mut Q-GDH</b> <b>Mut Q-GDH</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>Mut Q-GDH</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b> <b>Mut Q-GDH</b> <b>Mut Q-GDH</b> <b>Mut Q-GDH</b> <b>Mut Q-GDH</b> <b>Mut Q-GDH</b> <b>Mut Q-GDH</b> <b>Mut Q-GDH</b> <b>Mut Q-GDH</b> <b>Mut Q-GDH</b> <b>GDH-FAD</b>	<b>Roche Diagnostics</b> <a href="http://www.accu-chek.com">www.accu-chek.com</a> <b>Phone: +356 2134 4610</b>
<b>TRUE METRIX</b> <b>TRUE METRIX GO</b> <b>TRUE METRIX Air</b> <b>TRUeresult<sup>3</sup></b> <b>TRUeresult Twist 2<sup>3</sup></b> <b>TRUeyou<sup>3</sup></b>	<b>Yes</b> <b>Yes</b> <b>Yes</b>  <b>Not Recommended</b> <b>Not Recommended</b> <b>Not Recommended</b>	<b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b>  <b>GDH-FAD</b> <b>GDH-FAD</b> <b>GDH-FAD</b>	<b>Trividia Health</b> <a href="http://www.trividiahealth.com">www.trividiahealth.com</a> <b>Phone: N/A</b>

Glucose Monitor Brand	Compatible with Extraneal (Icodextrin) PD solution (Glucose-specific)	Test Type*	Manufacturer
Rightest Blood Glucose Monitoring System GM700	Yes	GDH-FAD	Bionime Corporation <a href="http://www.bionime.com">www.bionime.com</a> Phone: N/A
Rightest Blood Glucose Monitoring System GM700 Pro	Yes	GDH-FAD	
Rightest Blood Glucose Monitoring System GM700S	Yes	GDH-FAD	
Rightest Blood Glucose Monitoring System GM700SB	Yes	GDH-FAD	
Rightest Blood Glucose Monitoring System GM700 Pro 2	Yes	GDH-FAD	
Rightest Blood Glucose Monitoring System GM720	Yes	GDH-FAD	
Rightest Blood Glucose Monitoring System GM722	Yes	GDH-FAD	
Rightest Blood Glucose Monitoring System Max	Yes	GDH-FAD	
Rightest Blood Glucose Monitoring System Max Plus	Yes	GDH-FAD	
Rightest Blood Glucose Monitoring System Wiz	Yes	GO	
Rightest Blood Glucose Monitoring System Wiz Plus	Yes	GO	
Rightest Blood Glucose Monitoring System GM550	Yes	GO	
Rightest Blood Glucose Monitoring System GM280	Yes	GO	
Rightest Blood Glucose Monitoring System GM280B	Yes	GO	
Rightest Blood Glucose Monitoring System GT333	Yes	GO	
Rightest GT333 Blood Glucose Monitoring System	Yes	GO	
Rightest Blood Glucose Monitoring System GM300	Yes	GO	
Rightest Blood Glucose Monitoring System GM250	Yes	GO	
Rightest Blood Glucose Monitoring System GM260	Yes	GO	
GE100 Blood Glucose Monitoring System	Yes	GO	
Diathrive+ Blood Glucose Monitoring System	Yes	GO	

Glucose Monitor Brand	Compatible with Extraneal (Icodextrin) PD solution (Glucose-specific)	Test Type*	Manufacturer
Diathrive+ Bluetooth Blood Glucose Monitoring System	Yes	GO	
iGlucose Blood Glucose Monitoring System	Yes	GO	
GE333 Blood Glucose Monitoring System	Yes	GO	
Quintet AC Blood Glucose Monitoring System	Yes	GO	
Mylife Pura Blood Glucose Monitoring System	Yes	GO	
Mylife Unio Blood Glucose Monitoring System	Yes	GDH-FAD	
Mylife Unio Neva Blood Glucose Monitoring System	Yes	GDH-FAD	
Mylife Unio Cara Blood Glucose Monitoring System	Yes	GDH-FAD	
LineaD SIRIO Blood Glucose Monitoring System	Yes	GO	
LineaD EVO Blood Glucose Monitoring System	Yes	GO	
LineaD ORO HR Blood Glucose Monitoring System	Yes	GDH-FAD	
LineaD ORO Blood Glucose Monitoring System	Yes	GDH-FAD	
LineaD ORO24 Blood Glucose Monitoring System	Yes	GDH-FAD	
Rightest Blood Glucose Monitoring System Max Tel	Yes	GDH-FAD	
Rightest Blood Glucose and Ketone Monitoring System Max Dual	Yes	GDH-FAD	

<sup>1</sup> These Abbott monitors (together with the Dexcom G4 Platinum, G5 Mobile, G6 and the Medtronic iPro2) are the leading Continuous Glucose Monitor Systems (CGM) and are **not recommended** by their manufacturers for use into dialysis patients, and despite having a GO based chemistry currently there is not definitive scientific evidence for their Extraneal compatibility. Consult manufacturer for additional information.

<sup>2</sup> The ACCU-CHEK Nano (not Aviva or Performa) and ACCU-CHEK Aviva Plus glucose systems are available within the United States ONLY and use test strips branded as ACCU-CHEK SmartView and ACCU-CHEK Aviva Plus, respectively. These systems use the Mut Q-GDH (**Compatible**) strips. Consult manufacturer for additional information.

<sup>3</sup> TRUEresult (GDH-FAD), TRUEresult Twist 2, TRUEyou, were not tested with the Baxter recommended limits for Icodextrin and Maltose. So, these monitors **are not recommended** to use for dialysis patients.

**Test Type\***

**GO** = Glucose Oxidase

**GDH-NAD** = Glucose dehydrogenase with nicotinamide-adenine dinucleotide

**GDH-FAD** = Glucose dehydrogenase with flavin-adenine dinucleotide

**Mut Q-GDH** = Glucose dehydrogenase with pyrroloquinolinequinone modified to eliminate maltose interference

Please see full prescribing information for EXTRANEAL (Icodextrin) PD solution.

Baxter and EXTRANEAL are trademarks of Baxter International Inc.

Other trademarks or products appearing herein are the property of their respective owners.