

WARNING

Potential for Incorrect Blood Glucose Reading

ONLY use glucose monitors and test strips that are glucose-specific. These methods are common in clinical laboratories. Contact the manufacturer of the glucose monitors and test strips to determine the method that is used.

DO NOT use glucose monitors or test strips that utilize glucose dehydrogenase pyrroloquinolinequinone (GDH-PQQ) or glucose-dye-oxidoreductase methods. In addition, some but not all monitors or test strips that utilize a glucose dehydrogenase flavin-adenine dinucleotide (GDH-FAD) method should not be used. Use of these methods may result in falsely elevated blood glucose readings in patients using **EXTRANEAL** (icodextrin) due to maltose interference. Falsely elevated blood glucose readings may mask true hypoglycemia or lead to the erroneous diagnosis of hyperglycemia, leading to life-threatening events

Visit www.glucosafety.com for additional information, including a glucose monitor compatibility list."

Vantive

See reverse side for warnings

Patient Name

Is using **EXTRANEAL** (icodextrin)
peritoneal dialysis solution

Emergency Contact Information

Dialysis Unit _____

Other Contact _____

Vantive and Extraneal are trademarks of Vantive Health LLC or its affiliates.
Date of Preparation Aug 2024.